

Quality of Radon Measurements

- Radon and Radon Progeny Chambers for Calibration -

Martin Dubsclaff
14.02.2019, EURADOS winter school 2019
Łódź, Poland



Content

Radon measurements

Radon measurands
Examples of instruments

Calibration prerequisites

Source
Chamber
Measurement standard

Calibration procedures - Rn-222

Example laboratory: BfS
Traceability
Calibrations
Interlaboratory comparisons


Calibration procedures - Rn-222 Dp

Summary

2



Radon measurands – Rn-222 gas



Rn-222 activity concentration


- average amount of Rn-222 decays per second within a volume
- measurement unit: [**Bq/m³**]

Rn-222 exposure

- Rn-222 activity concentration multiplied by exposure time
- measurement unit: [**Bq·h/m³**]
- measuring instruments for this measurand: “**Exposimeters**“

3

Radon measurands – Rn-222 decay products



Potential Alpha-Energy

$$\text{PAE}_{\text{Rn-222}} [\mu\text{J}] \approx N_{\text{Po-218}} \cdot E_{\alpha, \text{Po-218}} + (N_{\text{Po-218}} + N_{\text{Pb-214}} + N_{\text{Bi-214}} + N_{\text{Po-214}}) \cdot E_{\alpha, \text{Po-214}}$$

with N_i : number of atoms of nuclide i ; α -energies: $E_{\alpha, \text{Po-218}} = 0.962 \text{ pJ}$, $E_{\alpha, \text{Po-214}} = 1.232 \text{ pJ}$

Potential Alpha-Energy Concentration → PAE per volume

$$\text{PAEC}_{\text{Rn-222}} [\mu\text{J}/\text{m}^3] = 5.56 \text{ nJ/Bq} \cdot C_{\text{Rn-222,eq}} [\text{Bq}/\text{m}^3] \quad \text{in case of radioactive equilibrium}$$

with $C_{\text{Rn,eq}}$: Rn-222 activity concentration in equilibrium with short-lived decay products

Equilibrium Equivalent (activity) Concentration

→ a radon activity concentration, in equilibrium with its short-lived decay products, which would have the same PAEC as the actual existing mixture of nuclides

$$\text{EEC}_{\text{Rn-222}} [\text{Bq}/\text{m}^3] = 0.105 \cdot C_{\text{Po-218}} + 0.516 \cdot C_{\text{Pb-214}} + 0.379 \cdot C_{\text{Bi-214}} + 10^{-8} \cdot C_{\text{Po-214}}$$

with C_i : activity concentration of nuclide i
4

Examples of radon measuring devices

Exposimeters (integrating detectors)
measurand: exposure [Bq*h/m³]

- solid-state nuclear track detectors (SSNTD)
- electrets



5



Examples of radon measuring devices

Electronical measuring instruments, measurand: Rn-222 activity concentration [Bq/m³]

- **scintillation cells**

e.g., flow-through scintillation cells with photo multiplier + analyzer and controlled flow



- **alpha spectrometry (filtered air, Si detector)**



- **ionisation chambers**



picture: Handbuch Radona Expert+ Waltec GmbH

6




Examples of radon PAEC measuring devices

Electronical measuring instruments – measurand, e.g., PAEC [$\mu\text{J}/\text{m}^3$]

- alpha spectrometry with filter



- with filter + wire screen (unattached fraction f_p)



7



Calibration

Calibration (ISO/IEC Guide 99:2007)

Operation that, under **specified conditions** [...] establishes a relation between the **quantity values** with measurement **uncertainties** provided by measurement standards [e.g., reference measurement] and [...] uses this information to **establish a relation** [e.g., calibration factor or curve] for obtaining a measurement result from an indication [e.g., of a **calibration object**].

e.g., an instrument for the measurand activity concentration showing indication value C_M



e.g., a measurement standard for activity concentration C_{Ref} traceable to national standards



$$C_M \neq C_{Ref}$$

→ **main goal**: find a relation $k(C_{Ref}, C_M)$ to correct indication values and consider measurement **uncertainties**

→ **adjustment** of a calibration object is **not included**
(a new calibration is necessary if the object has been modified)

8

Calibration prerequisites – Source, Chamber, Measurement standard

flow-through sources



for example
made by CMI (CZ)

emanator:
dry Ra-226 (salt)

activity: 500 kBq

purged with air

picture: [www.eurostandard.cz/
Eurostandard-catalog-2011.pdf](http://www.eurostandard.cz/Eurostandard-catalog-2011.pdf)


gas standard



for example made by
LNHB (France)

certified Rn-222
activity with 50 kBq

other emanation sources



for example made by
PTB (Germany)

e.g., produced by
electro-deposition

9



Calibration prerequisites – Source, **Chamber**, Measurement standard

walk-in calibration chambers

- practical issues
 - one can go inside to fix problems / look after instruments
 - **more/larger devices** can be exposed
- bigger volume means **less disturbance** of
 - radon activity concentration, e.g., while inserting objects
 - **surface influences** for Rn-222 decay products
 - PAEC measurements only possible in larger chambers**
(but not too large for homogeneity reasons)
- often adjustment of **environmental parameters** possible
- expensive, need space



small calibration chambers (containers)

- smaller sources / **less Rn-222 activity** needed
- good for operation of **many containers in parallel**
e.g., at different activity concentrations each
- good for long-term exposures without blocking a whole walk-in chamber




10

Calibration prerequisites – Source, **Chamber**, Measurement standard

Modes of chamber operation

decay mode

- one activity injection, e.g., gas standard
- decay of activity concentration $C_{Rn}(t)$
- e-Function, Rn-222 decay constant λ_{Rn} is known
- reference activity concentration can be
 - a) **measured** or
 - b) **calculated** if chamber volume and the initial activity is known



constant mode

- decay of activity concentration $C_{Rn}(t)$
- redosing needed to compensate decay
- reference activity concentration is **measured**
- constant mode needed for long term exposures, e.g., **integrating detectors** and for **Rn decay product calibrations**



11

Calibration prerequisites – Source, Chamber, **Measurement standard**



e.g., activity standard
+ volume standard
= activity concentration

National (measurement) standard


*Measurement standard **recognized by national authority** to serve in a state [...] as basis for assigning quantity values to other measurement standards [...].*

Reference (measurement) standard

*Measurement standard **designated for the calibration** of other measurement standards [...] in a given organization or at a given location.*

Working (measurement) standard

*Measurement standard that is **used routinely to calibrate** or verify measuring instruments or measuring systems.*



definitions acc. to ISO/IEC Guide 99:2007

12



Calibration procedures – BfS radon calibration laboratory

- **ISO 17025** accredited since 1999 as calibration laboratory
- accredited measurands **Rn-222 activity concentration** and **PAEC** of short-lived Rn-222 decay products
- non-accredited: Rn-220 (thoron) activity concentration



- 2 walk-in chambers: **30 m³** and **11 m³**
6 stainless steel containers: **0.4 m³** each
1 vacuum chamber **0.17 m³**
- accredited ranges
activity concentration **50 ... 12 000 Bq/m³**
PAEC **0.32 ... 640 μJ/m³**
- mainly **flow-through Ra-226** sources (CMI, Pylon)
- control of **temperature, humidity, pressure, turbulence**, amount and size spectra of **aerosols**

13



Realization of the measurand Rn-222 activity concentration

certified Rn-222 activity + certified volume = Rn-222 activity concentration

PTB Physikalisch-Technische Bundesanstalt
Braunschweig und Berlin
Nationales Metrologieinstitut

Seite 2 zum Kalibrierschein vom 12.12.2016, Kalibrierzeichen: PTB-6.13-11/2016
Page 2 of calibration certificate of 12.12.2016, calibration mark: PTB-6.13-11/2016

Gegenstand: Object	Aktivitätsnormal Activity standard
Typ: Type	Edelstahlzylinder mit Rn-222 High-grade steel cylinder containing Rn-222
Kenn-Nummer: Serial number	2016-1537
Radionuklid: Radionuclide	Radon-222 Radon-222
Kalibrierverfahren: Method of calibration	Die Aktivität des Rn-222-Gases wurde durch Messung der Gammastrahlung der kurzlebigen, mit Rn-222 im radioaktiven Gleichgewicht stehenden Folgeprodukte mit einem kalibrierten NaI(Tl)-Detektor bestimmt. Die Kalibrierkonstante für diesen Detektor wurde mit Rn-222 Aktivitätsnormalen der PTB ermittelt, deren Aktivitätsbestimmungen mit einem Halbleiter-Alpha-Spektrometer mit bekanntem Raumwinkel erfolgten. <i>The activity of the Rn-222 gas was determined with a calibrated NaI(Tl) detector by measuring the gamma radiation emitted by the short-lived decay products, which were in radioactive equilibrium with Rn-222 at the time of measurement. The calibration constant of this detector was determined with Rn-222 standards. The activities of these standards were measured at the PTB with a semiconductor alpha spectrometer with defined solid angle.</i>
Aktivität: Activity	$A = (6,59 \pm 0,12) \text{ kBq}$
Bezugszeitpunkt: Reference date	21. September 2016, 00:00 Uhr MEZ 00:00 CET on 21 September 2016

vacuum chamber
empty: 168.1 dm^3
+ AG: 166.2 dm^3

activity transfer
(by vacuum)




pressure and
temperature
controlled



14



Realization of the measurand Rn-222 activity concentration



Calibration procedures - Rn-222 activity concentration

reference standard



calibration
of working standards

- scintillation cells and
- flow-through scintillation cells
- Alphaguards, Atmos



16





Calibration procedures - Rn-222 activity concentration



Example: BfS radon calibration chamber (11 m³)

working standard: flow-through scintillation cell calibrated with reference standard

Pylon flow-through Ra-226 source with flow separation to adjust redosing





- temperature
- rel. humidity
- atmospheric pressure

17



Calibration procedures - Rn-222 activity concentration




→ “Operation [...] under **specified conditions** ...”

→ **stable** concentration mode (continuous dosing)

18

Calibration procedures - Rn-222 activity concentration

Calculation of the calibration factor k_M



$$k_M = \frac{C_{Ref}}{C_M - C_0}$$

mean Rn-222 activity concentration within Δt

C_M : of the calibration object

C_{Ref} : of the working standard

C_0 : blank indication of the calibration object

Δt depends on calibration object's sensitivity

Radon-222 activity concentration

$$C_0 = 4 \text{ Bq/m}^3$$

$$U(C_0) = 5 \text{ Bq/m}^3$$

calibration results

Nr.*)	$C_M - C_0$	$U(C_M - C_0)$	C_{Ref}	$U(C_{Ref})$	k_M	$U(k_M)$	Δt	T	$r.H.$	p
	Bq/m ³	Bq/m ³	Bq/m ³	Bq/m ³			Stunden/ Hours	°C	%	hPa
1	320	22	272	27	0,85	0,11	54	25	49	1003
2	1622	98	1363	118	0,84	0,09	4	26	38	1001
3	4712	286	3489	243	0,74	0,08	4	25	43	1010

*) Nummer der Messung / Number of measurement

example table (not the same calibration as in the graph)

19



Calibration procedures - Rn-222 activity concentration

Accreditation certificate D-K-15063-01-00 (BfS)

measurand	measurement range	minimal reportable uncertainty U(k=2)
Rn-222 activity concentration	50 – <80 Bq/m ³	17 %
	80 – <300 Bq/m ³	12 %
	300 – <1000 Bq/m ³	8 %
	1000 – <12000 Bq/m ³	5 %
PAEC of short-lived Rn-222 decay products	0.32 – <6.4 μJ/m ³	10 %
	6.4 – <64 μJ/m ³	6 %
	64 – <640 μJ/m ³	6 %

If uncertainty calculations show higher uncertainty values than these are reported.

calibration object

- st. error of the mean μ
- std. uncertainty of blank indication

$$u(\mu) = \frac{SD}{\sqrt{N}}$$

working standard

- std. error of the mean μ
- std. uncertainty of blank indication
- previous calibration hierarchy and drift

expanded uncertainty U = k * u with coverage probability >95%, k=2 for large number of values N (see also GUM JCGM 100:2008 Table G.2, e.g., k = 2.57 for N = 6)

20



Calibration procedures - Rn-222 activity concentration

Bundesamt für Strahlenschutz
 Radon-Kalibrierlaboratorium
 Köpenicker Allee 120–130
 10318 Berlin



Bundesamt für Strahlenschutz

akkreditiert durch die / accredited by the

Deutsche Akkreditierungsstelle GmbH



Deutsche
 Akkreditierungsstelle
 D-K-15063-01-00

als Kalibrierlaboratorium im / as calibration laboratory in the

Deutschen Kalibrierdienst



Kalibrierschein
 Calibration certificate

Kalibrierzeichen
 Calibration mark

1000
D-K- 15063-01-00
2018-05

Gegenstand
 Object

Radonmessgerät / Radon measuring instrument

Hersteller
 Manufacturer

Super example company GmbH/Ltd.

Typ
 Type

VeryBestRadonDevice 2000

Fabrikat/Serien-Nr.
 Serial number

12345

Dieser Kalibrierschein dokumentiert die Rückführung auf nationale Normale zur Darstellung der Einheiten in Übereinstimmung mit dem Internationalen Einheitensystem (SI).

This calibration certificate documents the traceability to national standards, which realize the units of measurement according to the International System of Units (SI).

Die DAkkS ist Unterzeichner der multilateralen Übereinkommen der European co-operation for Accreditation (EA) und der International Laboratory Accreditation Cooperation (ILAC) zur ge-

21



Interlaboratory comparisons and proficiency testing

Messgeräte zur Bestimmung der Radon-222-Aktivitätskonzentration oder der Radon-222-Exposition Vergleichsprüfung 2018

Instruments to Measure Radon-222 Activity Concentration or Exposure to Radon-222 - Interlaboratory comparison 2018

Fachbereich
Strahlenschutz und Umwelt

see full reports at <http://doris.bfs.de/jspui/>

(search for "interlaboratory comparison")


Elisabeth Foerster
Felice Friedrich
Martin Dubsloff
Felix Schneider
Joachim Döring



- 35 SSNTD or 24 electrets per participant
- 4 exposure groups for SSNTD (3 for electrets)
- at least 2 exposure values close together
- results anonymized



Referenzexposition /
Reference Exposure : 203 kBq h / m³



22



Interlaboratory comparisons and proficiency testing

Additional **proficiency testing** since 2019 due to new Radiation Protection Ordinance with relative upper (UL) and lower (LL) **performance limits** shown in the trumpet curve:

Test with data from BfS Interlaboratory Comparison 2018

Boxplots and outliers of all 4 exposure groups


Group 1: **203** kBq*h/m³

Group 2: **221** kBq*h/m³

Group 3: **1629** kBq*h/m³

Group 4: **1676** kBq*h/m³

● outliers of proficiency testing



23

Interlaboratory comparisons and proficiency testing

Outliers to pass the test

number of detectors	accepted number of outliers
8 - 17	0
18 - 27	1
28 - 37	2
38 - 47	3

derived from:

Beck T. R., Foerster E., et al., *The measurement accuracy of passive radon instruments*, Rad. Prot. Dosim. (2014), Vol. 158, No. 1, S. 59-67

To pass the test is one requirement for institutions that want to provide exposimeters for radon measurements in Germany acc. to the new Radiation Protection Ordinance.

24

Interlaboratory comparisons

Blank indication interlaboratory comparison 2017 of BfS and ENEA

- Deliverable D1.6.7 of MetroERM, JRP EMRP-ENV57
- Project coordinator S. Neumaier (PTB) as independent referee
- AlphaGuard from ENEA sent to BfS, reports sent directly to PTB only
- Measurement of several days during radon-free gas purging

Results (all uncertainties with $k=1$)

ENEA-INMRI / Italy, Francesco Cardellini:

$(2.5 \pm 0.5) \text{ Bq/m}^3$

BfS Radon calibration laboratory / Germany:

$(2.9 \pm 0.6) \text{ Bq/m}^3$

MetroERM
radiological
early warning



Also for measurements $< 300 \text{ Bq/m}^3$ (EU BSS) blank indication gets more important!

Interlaboratory Comparison of Rn measurement devices

- work in progress: MetroRadon WP 5.2.4.
- BfS sent device to different metrological institutions
- calibration at 400 Bq/m^3 , 1 kBq/m^3 and 6 kBq/m^3

Metro
RADON



The EMPIR initiative is co-funded by the European Union's Horizon 2020 research and innovation programme and the EMPIR Participating States


→ see also presentation by
Hannah Wiedner, BEV

25



Calibration procedures - Rn-222 decay products (PAEC)

Example: **BfS PAEC chamber (30 m³)** and calibration procedure



aerosols by
NaCl atomizer

aerosol counting
and spectra with
CPC & SMPS

26



Calibration procedures - Rn-222 decay products (PAEC)

sampling rod with glass fiber filter + wire screen on top


→ collection of decay products (attached and unattached)



27



Calibration procedures - Rn-222 decay products (PAEC)



- determined **filter efficiency** (>99.9%)
- wire screen is used to determine unattached fraction only
- **alpha spectrometry**
 - calibrated via combined or sequential α and γ measurement (α from Po-214 and γ from Bi-214)
 - Po-214 and Bi-214 are always in equilibrium

traceability:


Ra-226 plane source (Dp in equilibrium) with PTB certified activity and **calibrated volume flow rate**



28

Radon chambers in Europe that offer calibration services

(as far as known to the author)



29



Summary

- **measurands** Rn-222 activity concentration and PAEC
- examples of radon **measuring instruments**
- **sources**, characteristics of **chambers** and measurement **standards**
- calibration **procedures** with examples of the BfS radon calibration laboratory
- interlaboratory **comparisons** and overview over radon calibration services

Acknowledgements to ...

F. Friedrich, T. Beck, E. Foerster, S. Feige
Section UR1 “Radon Metrology“, BfS Berlin

P. Bossew
Section UR2 “Radon and NORM“, BfS Berlin

F. Wissmann
Division UR “Environmental Radioactivity“, BfS Berlin

... and **YOU** for your attention.

contact:

mdubslaff@bfs.de

